

Response Factors

In the ideal case, a Flame Ionization Detector (FID) is a carbon ion counter, whose response is directly proportional to the carbon count of the material being analyzed. The response is generally output as a concentration of carbon atoms in the medium being sampled, such as "parts per million carbon as vapor" or, for example ppm C (carbon). However physical detector cell design limitations, the physics of the ion stream in the hydrogen flame and the molecular structure and composition of the gas will affect the response. The FID output in many cases will NOT be equal to the actual carbon count. It will however be repeatable and proportional. For example, a FID calibrated with 1000 ppm propane (C_3H_8) will read 3000 ppm C_1 . If 3000 ppm methane is sampled, the response should be the same, i.e. 3000 ppm C_1 . In actual practice the reading is higher, but will always be higher by the same factor. This factor is called the Response Factor (RF). The RF phenomenon is well known. There can be RF for every compound. In some cases it will be unity ("1"). Ideally, one should know the RF for all compounds being measured, so that the necessary corrections may be applied. If possible, one should calibrate the analyzer (FID) with a known gas of the material being sampled. This is usually not practical, especially in cases where there are many gas constituents, or the species are unknown, as in combustion products or other emissions.

In practice, most Total Hydrocarbon Analyzers with FID's are calibrated with propane span gases, and the results are reported as ppm C_3 , (propane equivalent). If the response factor is known, the results may be converted to true concentrations, or the Total Hydrocarbon Analyzer (THA) may be initially calibrated with the response factor built in.

The following formula defines the calibration in mgC/m^3 :

$$f_{RCx} = \frac{A_{Cx}}{C_{Cx}}$$

f_{RCx} : carbon responsefactor of the component x, propane equivalent

A_{Cx} : measured (displayed) value of the component x in mgC/m^3 ; FID calibrated with propane, with 1ppm $C_3H_8 = 1,61 mgC/m^3$ under standard conditions (273 K, 1013 hPa)

C_{Cx} : carbon concentration in mgC/m^3 of the measured component x

To convert the measured (displayed) value into the actual concentration:

$$C_{Cx} = \frac{A_{Cx}}{f_{RCx}}$$

For the calibration in ppm the numerical values of f_{RCx} are the same but the formula has to be changed as follows:

$$f_{RCx} = \frac{[A_{Vx} \cdot n_{Ref}]}{[C_{Vx} \cdot n_x]}$$

f_{RCx} : carbon responsefactor of the component x, propane equivalent

A_{Vx} : measured (displayed) value of the component x in ppm; FID calibrated in ppm propane

C_{Vx} : volume concentration of the measured component x in ppm

n_{Ref} : carbon number of the reference molecule; for propane: C_3H_8 , 3 C-atoms: $n_{Ref} = 3$

n_x : carbon number of the substance x

To convert the measured (displayed) value into the actual concentration:

$$C_{Vx} = \frac{[A_{Vx} \cdot n_{Ref}]}{[f_{RCx} \cdot n_x]}$$

Technical Information

The following response factors (propane equivalent) were determined by TÜV-Nord using two J.U.M. FID's, Model VE 7, the FID 2 was tested 9 months after the FID 1

Component	Total Formula	f _{RCx} FID 1	f _{RCx} FID 2
Acetone	C ₃ H ₆	0,72	0,73
Acetylene	C ₂ H ₂	0,92	0,94
Benzene	C ₆ H ₆	1,05	1,05
n-Butane	C ₄ H ₁₀	0,98	0,98
Chlorobenzene	C ₆ H ₅ _Cl	1,01	1,04
Chloroform	HCCl ₃	0,82	0,78
Cyclohexane	C ₆ H ₁₂	0,93	0,94
Methylene Chloride	CH ₂ Cl ₂	1,09	1,06
Diethyl Ether	C ₄ H ₁₀ O	0,75	0,77
Acetic Acid	C ₂ H ₄ O ₂	0,58	0,55
Ethyl Acetate	C ₄ H ₈ O ₂ (H ₃ C_COOC ₂ H ₅)	0,70	0,72
Isobutyl Acetate		0,88	0,89
Ethanol	C ₂ H ₆ O	0,65	0,67
4-Ethyl Toluene		0,88	0,89
n-Heptane	C ₇ H ₁₆	0,91	0,95
n-Hexane	C ₆ H ₁₄	0,85	0,86
Methane	CH ₄	1,26	1,25
Methanol	CH ₄ O	0,69	0,68
iso-Octane		0,99	0,98
Propane	C ₃ H ₈	1,00	1,00
iso-Propanol	C ₃ H ₈ O	0,82	0,81
Tetrachlorethylene	C ₂ Cl ₄	1,22	1,20
Toluene	C ₇ H ₈	1,02	1,03
Trichlorethene	C ₂ HCl ₃	1,03	1,01
1,1,1-Trichlorethane	C ₂ H ₃ Cl ₃	1,06	1,02
p-Xyloene	C ₈ H ₁₀	0,91	0,90

Other C₃H₈ equiv. Response factors found with J.U.M. FID's VE7, 3-300A and 3-200

Component	Total Formula	f _{RCx}
Acetonitrile	C ₂ H ₃ N	0,7
2-Butanol	C ₄ H ₁₀ O	0,83
2-Butanone	C ₄ H ₈ O	0,78
<i>Chlorpropionsäuremethylester</i>		0,69
Dichlormethane		1,06
Diisopropylether	C ₆ H ₁₄ O	0,84
Dimethylformamide	C ₃ H ₇ NO	0,50
Dioxane	C ₄ H ₈ O ₂	0,51
Ethylbenzene	C ₆ H ₅ _C ₂ H ₅	1,04
Ethan	C ₂ H ₆	1,01
Ethin		1,07
<i>Etylenglykolmonoethylester</i>		0,58
Iso-Octane		0,98
Isopropylalcohol (Isopropanol)		0,79
Methanol	CH ₃ OH, CH ₄ O	0,79
Methylcyclohexane	C ₇ H ₁₄	0,96
Methylenechloride	CH ₂ Cl ₂	1,1
4-Methyl-2-Pentanone	C ₆ H ₁₂ O	0,83
Pentane	C ₅ H ₁₂	0,98
i-Propanol	C ₃ H ₈ O	0,74
n-Propanol	C ₃ H ₈ O	0,80
Propylchloride	C ₃ H ₇ Cl	1,05
o-Xyloene	C ₈ H ₁₀	1,01
Trichlormethane (Chloroform)		0,8
Trichloethene		1,07